



U.S. Department  
of Transportation

Pipeline and  
Hazardous Materials  
Safety Administration

East Building, PHH-23  
1200 New Jersey Ave, SE  
Washington, D.C. 20590

IAEA CERTIFICATE OF COMPETENT AUTHORITY  
SPECIAL FORM RADIOACTIVE MATERIALS

CERTIFICATE USA/0245/S-96, REVISION 14

This certifies that the sources described have been demonstrated to meet the regulatory requirements for special form radioactive material as prescribed in the regulations of the International Atomic Energy Agency<sup>1</sup> and the United States of America<sup>2</sup> for the transport of radioactive material.

1. Source Identification - Elekta AB Models 43047, 43685 and 1000029.
2. Source Description - Cylindrical double encapsulations made of Type 303 stainless steel (Model 43047) or Type 316L stainless steel (Models 43685 and 1000029) and seal welded. Approximate outer dimensions are either 7.98 mm (0.314 in.) or 7.99 mm (0.315 in.) in diameter and 27.2 mm (1.07 in.) in length. Construction shall be in accordance with attached Elekta Instrument Drawing Nos. 43047 (Model 43047), 43685 (Model 43685), or 1000029 (Model 1000029).
3. Radioactive Contents - No more than 2.22 TBq (60.0 Ci) of Cobalt-60 in the form of metal pellets, a solid rod, or powder.
4. Management System Activities - Records of Management System activities required by Paragraph 306 of the IAEA regulations shall be maintained and made available to the authorized officials for at least three years after the last shipment authorized by this certificate. Consignors in the United States exporting shipments under this certificate shall satisfy the requirements of Subpart H of 10 CFR 71.
5. Expiration Date - This certificate expires on February 28, 2026. Previous editions which have not reached their expiration date may continue to be used.

---

<sup>1</sup> "Regulations for the Safe Transport of Radioactive Material, 2012 Edition, No. SSR-6" published by the International Atomic Energy Agency (IAEA), Vienna, Austria.

<sup>2</sup> Title 49, Code of Federal Regulations, Parts 100-199, United States of America.

**CERTIFICATE USA/0245/S-96, REVISION 14**


This certificate is issued in accordance with paragraph(s) 804 of the IAEA Regulations and Section 173.476 of Title 49 of the Code of Federal Regulations, in response to the January 12, 2021 petition by Elekta Instrument, Stockholm, Stockholm County, Sweden, and in consideration of other information on file in this Office.

Certified By:



February 10, 2021


(DATE)

 William Schoonover  
Associate Administrator for Hazardous  
Materials Safety

Revision 14 - Issued to extend the expiration date.

This document is the property of ELEKTA and may not be copied or used for any other purpose without the owners' written consent. Controversy will lead to prosecution.

No.	Revision	Date	Appd
1	Utseende justerat	8804	RL
2	Not tillk; se ÄM C0192	9436	JT
3	Not tolkas tillk. ÄM C0480	9547	KES
4	Pos 2 ritn. tillk. ÄM C1321	9745	BTHO
5	See change order C1836	9930	BON
6	DWG. numbers added. ÄM C2186	9951	RL
7	See ECO-0000104	0506	OLNI



Kapseln tolkas med tolerans  $\varnothing 8H7\pm 0$  efter svetsning för att passa i bussning vid montering.


The capsule to be checked after welding with a tool having hole tolerance  $\varnothing 8H7\pm 0$ .

1) Gastät svets utan tillsatser/ Gas tight weld without additives  
 Not/ Note: Teknisk specifikation/Technical Specification Art.No. 9-01863 0.010

Item	Qty	Description	Material	Dimension, Article no.	Weight kg	
Conditions where other not is specified						
Date		Welding		Radius Sharp edges	Tolerances	
84-04-17		Checked		Scale	Ref.	
Drawn		KS		5:1	908786	
ELEKTA INSTRUMENT		PREPARATKAPSEL SMST CAPSULE ASSEMBLY			Sheet	Cont.
					43047	

This document is the property of ELEKTA and may not be copied or used for any other purpose without the owners' written consent. Contravention will lead to prosecution.

No.	Revision	Date	Appd
1	Utseende justerat	8804	RL
2	Not tillk; SE ÅM C0192	9436	JT
3	Not tolkas tillk. ÅM C0480	9547	KES
4	See change order C1836	9938	BON
5	See change order C2040	9947	RL



Kapseln tolkas med tolerans  $\varnothing 8H7 \pm 0$  efter svetsning för att passa i bussning vid montering.  
 The capsule to be checked after welding with a tool having hole tolerance  $\varnothing 8H7 \pm 0$ .


- 1) Gastät svets utan tillsatser/ Gas tight weld without additives
  - 2) Pluglängd efter behov/ Plug length as required
- Not/ Note: Teknisk specifikation/Technical Specification Art.No. 9-01863

0.010

Item	Qty	Description	Material	Dimension, Article no.	Weight kg
Conditions where other not is specified		Surface	Welding	Radius Sharp edges	Tolerances
Date	Drawn	Checked	Scale	Ref.	
84-04-17	KS/RL	MS	5:1	908788	
ELEKTA INSTRUMENT		PREPARATKAPSEL SMST CAPSULE ASSEMBLY		Sheet	Cont.
				43685	

This document is the property of ELEKTA and may not be copied or used for any other purpose without the owners' written consent. Controvention will lead to prosecution.

No.	Revision	Date	Appd
02	See ECO-231	0543	AB



Kapseln tolkas med tolerans  $\varnothing 8H7 \pm 0$  efter svetsning för att passa i bussning vid montering.  
 The capsule to be checked after welding with a tool having hole tolerance  $\varnothing 8H7 \pm 0$ .

- 1) Gastät svets utan tillsatser/ Gas tight weld without additives  
 2) Pluglängd efter behov/ Plug length as required

Not/ Note: Teknisk specifikation/Technical Specification Art.No. 9-01863

0.010

Item	Qty	Description	Material	Dimension, Article no.	Weight kg	
Conditions where other not is specified		Welding		Radius Sharp edges	Tolerances	
Date	Drawn	Checked	Scale	Ref.		
050701	OLNi			1000028		
ELEKTA INSTRUMENT		CAPSULE ASSEMBLY			Sheet	Cont.
					1000029	



U.S. Department of  
Transportation

**Pipeline and  
Hazardous Materials  
Safety Administration**

East Building, PHH-23  
1200 New Jersey Ave, SE  
Washington, D.C. 20590

**CERTIFICATE NUMBER:** USA/0245/S-96

**ORIGINAL REGISTRANT(S) :**

MDS Nordion  
MDS Nordion  
447 March Road  
Ottawa, Ontario, K2K 1X8  
Canada

GE Hitachi Nuclear Energy  
6705 Vallecitos Rd  
Sunol, CA, 94586  
USA

Elekta Instrument  
Box 7593  
SE-103 93  
Stockholm, Stockholm County,  
SWEDEN